

MAY 09 2002

CD HORIZON® Spinal System
Summary of Safety and Effectiveness
February 2002

I. Company: Medtronic Sofamor Danek, Inc. USA
1800 Pyramid Place
Memphis, TN 38132
(901) 396-3133

II. Proposed Proprietary Trade Name: CD HORIZON® Spinal System

III. Product Description

The CD HORIZON® Spinal System consists of a variety of rods, hooks, screws, CROSSLINK® plates, staples, and other connecting components used to build a spinal construct. Instrumentation is also available to facilitate implantation of the device components.

Certain implant components from other Medtronic Sofamor Danek spinal systems can be used with the CD HORIZON® Spinal System. These components include TSRH® rods, hooks, screws, plates, CROSSLINK® plates, connectors, staples and washers; GDLH® rods, hooks, connectors and CROSSLINK® bar and connectors; LIBERTY® rods and screws; DYNALOK PLUS® bolts; and Medtronic Sofamor Danek Multi-Axial rods and screws.

CD HORIZON® hooks are intended for posterior use only. CD HORIZON® staples and CD HORIZON® ECLIPSE® rods and screws are intended for anterior use only. However, for patients of smaller stature, CD HORIZON® 4.5mm rods and associated components may be used posteriorly.

The purpose of this 510(k) submission is to add titanium 4.5mm rods and associated components for posterior use for patients of smaller stature.

IV. Indications

When used in a percutaneous posterior approach with the SEXTANT instrumentation, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for the following indications:

When used as a posterior spine thoracic/lumbar system, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e. degenerative scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

In addition, when used as a pedicle screw fixation system the CD HORIZON® Cannulated M8 Multi-Axial Screw components are also indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); (d) who are having the device removed after the development of a solid fusion mass.

000047

The CD HORIZON® ECLIPSE® components are intended for the following indications:

When used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® system is also intended for the following indications:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the CD HORIZON® Spinal System is indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

In addition, when used as a pedicle screw fixation system, the CD HORIZON® Spinal System is indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

V. Substantial Equivalence

Documentation was provided which demonstrated the CD HORIZON® Spinal System to be substantially equivalent to itself.



DEPARTMENT OF HEALTH & HUMAN SERVICES

Public Health Service

MAY 09 2002

Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

Mr. Richard W. Treharne, PhD
Senior Vice President, Regulatory Affairs
Medtronic Sofamor Danek
1800 Pyramid Place
Memphis, Tennessee 38132

Re: K020709

Trade/Device Name: CD HORIZON® Spinal System
Regulation Number: 21 CFR §888.3050, §888.3060, §888.3070
Regulation Name: Spinal interlaminar fixation orthosis, Spinal intervertebral body
fixation orthosis and Pedicle screw spinal system
Regulatory Class: Class II
Product Code: KWP, KWQ, MNH, MNI
Dated: April 8, 2002
Received: April 9, 2002

Dear Dr. Treharne:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies.

You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 21 CFR Part 809.10 for *in vitro* diagnostic devices), please contact the Office of Compliance at (301) 594-4659. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,



for Celia M. Witten, Ph.D., M.D.
Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

510(k) Number (if known): K020709

Device Name: CD HORIZON® Spinal System

Indications for Use:

When used in a percutaneous posterior approach with the SEXTANT instrumentation, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for the following indications:

When used as a posterior spine thoracic/lumbar system, the CD HORIZON® Cannulated M8 Multi-Axial Screw components are intended for: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e. degenerative scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

In addition, when used as a pedicle screw fixation system the CD HORIZON® Cannulated M8 Multi-Axial Screw components are also indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (c) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); (d) who are having the device removed after the development of a solid fusion mass.

The CD HORIZON® ECLIPSE® components are intended for the following indications:

When used as an anterolateral thoracic/lumbar system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

The CD HORIZON® system is also intended for the following indications:

When used as a pedicle screw fixation system of the non-cervical posterior spine in skeletally mature patients, the CD HORIZON® Spinal System is indicated for one or more of the following: (1) degenerative spondylolisthesis with objective evidence of neurologic impairment, (2) fracture, (3) dislocation, (4) scoliosis, (5) kyphosis, (6) spinal tumor, and/or (7) failed previous fusion (pseudarthrosis).

000158

In addition, when used as a pedicle screw fixation system, the CD HORIZON® Spinal System is indicated for skeletally mature patients: (a) having severe spondylolisthesis (Grades 3 and 4) of the fifth lumbar-first sacral (L5-S1) vertebral joint; (b) who are receiving fusions using autogenous bone graft only; (3) who are having the device fixed or attached to the lumbar and sacral spine (L3 and below); and (4) who are having the device removed after the development of a solid fusion mass.

When used as a posterior, non-cervical, non-pedicle screw fixation system, the CD HORIZON® Spinal System is intended for the following indications: (1) degenerative disc disease (as defined by back pain of discogenic origin with degeneration of the disc confirmed by patient history and radiographic studies), (2) spinal stenosis, (3) spondylolisthesis, (4) spinal deformities (i.e., scoliosis, kyphosis, and/or lordosis), (5) fracture, (6) pseudarthrosis, (7) tumor resection, and/or (8) failed previous fusion.

(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF
NEEDED)

Concurrence of CDRH, Office of Evaluation (ODE)

Prescription Use X
(Per 21 CFR 801.109)

OR

Over-the-counter Use _____

(Optional 1-2-96)

Miriam C. Provost
(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K020709

00015